

Helica Biosystems, Inc
231 E. Imperial Hwy.
Ste. 250
Fullerton, CA 92835



helica
biosystems inc

Phone: 714-578-7830
Fax: 714-578-7831
Email: info@helica.com

Introducing Low Matrix ELISA for Aflatoxin and Ochratoxin A detection in difficult-to-assay commodities

Helica Biosystems, Inc is pleased to announce an innovative approach to testing aflatoxins and ochratoxin A in commodities which pose difficult problems in measurement by conventional “rapid” ELISA methods.

The so-called rapid ELISA method provides for a quicker and simpler assay with less manipulation by the operator and is used where speed and simplicity are at a premium, for instance when large truckloads of grain must be screened for mycotoxin content.

The main drawback with this type of assay is the well known ‘matrix effect’. These problems are compounded by the fact that a mycotoxin assay may be used with many different kinds of commodity extract, and controlling for matrix effect in one commodity by, for instance, using an antigen-free extract in the assay would only correct for the effects of that particular matrix and would likely be incorrect for others. A case in point is the assay of aflatoxin in silage, a vegetable matter usually a byproduct of grain harvests preserved by fermentation in a large enclosed structure or silo and is an important component of cattle fodder. Immunoassay of aflatoxin in silage extract in a conventional rapid assay leads to high false positive values in samples that contain no aflatoxin by the HPLC gold standard.

A similar situation exists in the case of ochratoxin A. While most grains and cereals, with a relatively moderate matrix effect, may be screened for ochratoxin content using the simultaneous incubation of sample and conjugate, commodities such as coffee, cocoa, spices, and grape derivatives such as red wine are notoriously difficult to test due to “matrix effect”.

As a solution to this problem, Helica Biosystems, Inc has successfully devised immunoassays to measure total aflatoxins and ochratoxin in silage, coffee and other high matrix effect commodities at biologically useful levels. **These low matrix assays are now available for distribution world-wide (see test list in Appendix).** We are pleased to present the following set of data from a study we conducted to validate the field utility of this method.

PART 1 – LOW MATRIX ELISA FOR AFLATOXINS

Studies comparing Low Matrix ELISA with the HPLC gold standard

Extraction study

Thirteen samples of finely ground silage: five corn silage (C1 – C5), two wheat silage (W1 – W2), three hay silage (H1 – H3) and three snaplage (S1 – S3) samples which had previously measured negative for aflatoxins B1 at the 1 ppb level by HPLC following solid-phase clean up of the extracted samples.** These samples were extracted by vigorous vortexing in either 80% methanol or 80% acetonitrile at a 1:5 ratio of commodity to solvent for two minutes and measured in the Helica Biosystems low matrix total aflatoxin ELISA (**Cat.no. 981AFL01LM**). Results are presented below.

Sample	Acetonitrile extract		Methanol extract	
	%B ₀	ppb	%B ₀	ppb
C1	90.2	<1	92.1	<1
C2	88.2	<1	92.6	<1
C3	91.5	<1	89.7	<1
C4	71.7	1.8	91.5	<1
C5	92.2	<1	90.1	<1
W1	78.2	1.3	79.6	1.2
W2	86.2	<1	84.3	<1
H1	74.2	1.6	83.1	<1
H2	78.6	1.3	84.5	<1
H3	76.6	1.5	89.3	<1
S1	99.5	<1	95.6	<1
S2	98.9	<1	95.6	<1
S3	91.1	<1	96.8	<1

Recovery Study

Each of the different types of silage was also spiked with 5 ng/gm (5ppb) of aflatoxin B1 and dried down overnight at room temperature in order to estimate the recovery of aflatoxin by the two different solvents. The following results were obtained:

Sample	Acetonitrile extract		Methanol extract	
	ppb	% Recovery	ppb	% Recovery
Spike	4.8	100	5.1	100
C1	4.1	85	2.5	49
W2	4.8	100	2.7	53
H1	4.6	96	2.7	53
S1	4.6	96	2.9	57

The following **extraction and recovery** experiments were also performed **on store-bought samples of paprika, pistachio and peanuts** which are commodities also prone to contamination by aflatoxin

Extraction Study

Sample	Acetonitrile extract		Methanol extract	
	%B ₀	ppb	%B ₀	ppb
Paprika	96.5	<1	99.2	<1
Pistachio	96.2	<1	96.9	<1
Peanut	99.6	<1	102.2	<1

Recovery Study

	Acetonitrile extract		Methanol extract	
	ppb	% Recovery	ppb	% Recovery
Spike	5.4	100	4.5	100
Paprika	5.2	96	3.0	67
Pistachio	5.0	93	3.1	69
peanut	3.6	67	2.6	58

Conclusion

While 70-80% methanol is frequently used in the extraction of commodities for the assay of aflatoxins by ELISA, it can clearly be seen from these results that acetonitrile extraction is more efficient than methanol, giving efficiencies of 85-100% in all cases except peanuts which was 67%. If methanol is used for extraction its extraction efficiency of 50-70% should be taken into account.

It is difficult to determine at this point whether the low (<2ppb) values measured in some silage samples represents residual matrix effect or the cumulative measurement of several aflatoxin types (B1,B2,G1,G2) present at less than 1 ppb levels. **Otherwise the Helica Biosystems low matrix ELISA is a cost effective screening tool for total aflatoxin content in a wide variety of commodities destined for animal or human consumption.**

*** We would like to acknowledge Trilogy Analytical Laboratory www.trilogylab.com (Washington, MO, USA) for kindly performing the HPLC analysis in this study.*

PART 2 – LOW MATRIX ELISA FOR OCHRATOXIN A

Ochratoxin A in Wine and Beer

Ochratoxin-free samples of grape products and beer were spiked with ochratoxin in levels from 0.4 to 8.0 ng/mL in wine and 0.04 to 0.8 ng/mL in the case of beer. After a simple 1:20 dilution in methanol (1:2 for beer), they were assayed for ochratoxin A with the Helica Biosystems quantitative assay for ochratoxin in alcoholic beverages (**Cat. no. 981OCH01ALC**). Recoveries of ochratoxin A from the various commodities compared to the kit standards are given below:

Standard	Red Wine%	White Wine%	Port %	Sherry %	Must %	Juice %	Beer %
0.02	93	85	100	105	98	65	120
0.05	104	98	102	104	101	72	112
0.10	100	92	98	105	99	80	115
0.20	103	110	98	93	98	93	120
0.40	110	99	108	101	91	95	113

The results demonstrate that the Helica Biosystems Quantitative Ochratoxin A assay can be used to measure Ochratoxin A in a wide variety of alcoholic and non-alcoholic beverages. Please see product package insert for more detailed data.

Ochratoxin A in Coffee, Cocoa, and Spices

Samples were spiked with 5 ug/Kg (ppb) of ochratoxin A and dried overnight. They were then extracted in 80% acetonitrile in a 5:1 ratio (solvent:commodity). Acetonitrile was shown to be superior to methanol for this purpose. After a 10-fold dilution in 70% methanol, the extracts were measured in the Helica Biosystems Quantitative assay for ochratoxin in coffee, cocoa, and spices (**Cat. no. 961OCH01COF**). Extraction and recovery experiments were performed three times for each commodity with the following results:

	Recovery 1 (%)	Recovery 2 (%)	Recovery 3(%)
Green Coffee	99	96	90
Roast Coffee	91	97	92
Instant Coffee	107	103	98
Cocoa Powder	110	108	108
Cocoa Butter	97	104	100
Paprika	102	104	109
Chili Powder	101	102	107

Please see product package insert for more detailed data.

Ochratoxin A in Serum and Milk

As a result of the report “Assessment of Dietary Intake of Ochratoxin A by the Population of EU Members States: Report fo Experts Participating in Task 3.2.7, Jan 2002”, it is apparent that a simple procedure for the accurate measurement of ochratoxin A in serum and milk would be of great benefit in assessing the bioburden of this mycotoxin in human populations.

Human and animal serum and milk were spiked with 0.2 ng/mL of cohtratoxin A. After a 4-fold dilution in methanol were assayed in the Helica Biosystems Quantitative assay for ochratoxin in milk and serum (**Cat no. 991OCH01MS**). Recoveries were repeated three times and compared to recovery from a similarly spiked phosphate buffered saline solution. Results are presented below.

	Recovery 1(%)	Recovery 2(%)	Recovery 3(%)	Recovery mean (%)
Human Serum	102	102	104	103
Pig Serum	100	94	96	97
Human Milk	96	110	95	100
Cow's Milk	114	116	113	114

Please see product package insert for more detailed data.

This assay should provide a useful adjunct to the measurement of ochratoxin A in individual commodities destined for human and animal consumption.

Please contact us for additional information on these and other assays. Package inserts are available by request. A complete list of our MycoMonitor® line of ELISA assays follows in the appendix.

Helica Biosystems, Inc
231 E. Imperial Hwy.
Ste. 250
Fullerton, CA 92835



Phone: 714-578-7830
Fax: 714-578-7831
Email: info@helica.com

APPENDIX

CAT. NO.	HELICA BIOSYSTEMS, INC .	SIZE
	MYCOTOXIN ELISA	
941AFL01M	Total Aflatoxins, 96 tests grains and cereals, nuts, animal feed	96-tests
981AFL01LM	Total Aflatoxins (LOW MATRIX), 96 tests <i>NEW</i> grains and cereals, silage, nuts, spices, animal feed	96-tests
961AFLM01M	Aflatoxin M1 , 96 tests milk and dairy products	96-tests
941DON01M	Deoxynivalenol (DON), 96 tests grains and cereals, animal feed	96-tests
941OCH01M	Ochratoxin FOR GRAINS, 96 tests grains and cereals, animal feed	96-tests
981OCH01W-96	Ochratoxin FOR WINE, Qualitative, 96 tests <i>NEW</i> wine, grapes juice, grape must	96-tests
981OCH01W-48	Ochratoxin FOR WINE, Qualitative, 48 tests <i>NEW</i> wine, grapes juice, grape must	48-tests
981OCH01ALC	Ochratoxin Quantitative.,alcoholic beverages,96 tests <i>NEW</i> wine, grapes juice, grape must, beer	96-tests
961OCH01COF	Ochratoxin A for coffee , cocoa, and spices <i>NEW</i> coffee , cocoa, cocoa butter and spices	96-tests
991OCH01MS	Ochratoxin A for milk and serum <i>NEW</i> human and animal milk, serum, and plasma	96-tests